


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AUK-AV-P-05*

S.N.: *562*

Date: *JULY 20 81*



*10 V / DIV
5 nsec / DIV
20 kHz*



*10 V / DIV
1.0 nsec / DIV
20 kHz*

a) Output signal amplitude:

0 TO 7.15 V

b) Pulse width:

0 TO 50 nsec

c) Rise time:

≤ 100 nsec

d) Fall time:

≤ 2 nsec

e) PRF:

0 TO 50 kHz

f) Jitter, stability:

OK

g) Prime power:

±15 ±1.0 V

275 mA (IMAX)