


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: *AVH-MPT-2*

S.N.: *565*

Date: *Jan 4 27 81*



a) Output signal amplitude:

> +60V

b) Pulse width:

> 2 nsec

c) Jitter, stability:

OK

d) Prime power:

+15 ± 1.0 VOLT

150 mA (max)

32 VOLTS/DIV

2.0 nsec/DIV.

400 kHz

** SPECIAL UNIT*

PW = 2 nsec

AT FWHM RATHER

THAN 10% RISE

POINT.