


IMPULSE GENERATOR

PERFORMANCE CHECK


Model: *AVH-5-1-05-N*

S.N.: *575*

Date: *JULY 27 81*



*1.0 NS/C DIV  
10 VOLTS/DIV  
400 KHZ*



*100 PSEC/DIV  
10 VOLTS/DIV  
400 KHZ*

a) Output signal amplitude:

*-10V*

b) Pulse width:

*≤ 135 PSEC.*

c) Jitter, stability:

*OK*

d) Prime power:

*+15 ± 1.0 V*

*200 mA (MAX.)*

