


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AVL-2-N-T-PS*

S.N.: *579*


Date: *AUG 24 81*



MIN TR



MID TR



MAX TR

*≈ 340 VOLTS/DIV
10 NSEC/DIV
2 KHZ
R. = 50Ω*

- a) Output signal amplitude:
0 TO - 350 V (50Ω)
- b) Pulse width:
0 TO 100 NSEC
- c) Rise time:
≈ 1 - 10 NSEC
- d) Fall time:
≈ 2 - 10 NSEC
- e) PRF:
0 TO 5 KHZ
- f) Jitter, stability:
OK
- g) Prime power:
*120/240 V,
50-60 HZ*