


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AUM-2-P-PS*

S.N.: *590*

Date: *SEPT 14 81*



1.0 NS/E/DIV
10 VOLTS/DIV
4.0 MHz

- a) Output signal amplitude:
0 TO +15 V
- b) Pulse width:
0 TO 2 NS/E
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

*110/220V * 50-60 Hz*

[Signature]

** THIS UNIT SHIPPED
CONNECTED FOR
220V OPERATION.*

[Signature]