


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: **AVG-1-N**

S.N.: **605**

Date: **SEPT 28 81**



a) Output signal amplitude:

3-250 VOLTS

b) Pulse width:

≤ 4 NSEC

c) Jitter, stability:

O.K.

d) Prime power:

+15 ± 1.0 VOLT

275 mA MAX.

100 VOLTS/DIV.

5 NSEC/DIV.

20 KHZ.

Handwritten signature