


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: *AV6-2-N*

S.N.: *606*

Date: *SEPT 28 81*



a) Output signal amplitude:

$\geq -120 \text{ V}$

b) Pulse width:

$\leq 2 \text{ NSFC}$

c) Jitter, stability:

OK

d) Prime power:

$\pm 15 \pm 1.0 \text{ V}$

220 mA (MAX)

100 VOLTS/DIV

2 NSFC/DIV

20 KHZ.