


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: *AV6-2-N*

S.N.: *607*

Date: *SEPT 21 81*



a) Output signal amplitude:

$\geq -120 \text{ V}$

b) Pulse width:

$\leq 2 \text{ nSEC}$

c) Jitter, stability:

OK

d) Prime power:

$+15 \pm 1.0 \text{ VOLTS}$

260 mA (MAX)

100 VOLTS/DIV

2 nSEC/DIV

20 kHz