


PULSE GENERATOR

PERFORMANCE CHECK

Model: **AVM-2-P-D-PS**

S.N.: **609**

Date: **SEPT 14 81**



- a) Output signal amplitude:
0 TO +15 V
- b) Pulse width:
0 TO 2 NS
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power: *

1.0 NS/DIV

10 VOLTS/DIV

4.0 MHz

110/220V, 50-60 Hz

**THIS UNIT SHIPPED
CONNECTED FOR
240V OPERATION.**