


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-1-21 (NEGATIVE CHANNEL)*

S.N.: *624*

Date: *SEPT 22 81*



- a) Output signal amplitude:
0.70 - 5V
- b) Pulse width:
0 TO 6 NS
- c) Rise time:
≤ 100 PS
- d) Fall time:
≤ 135 PS
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK.
- g) Prime power:

5V / DIV
1.0 NS / DIV
4.0 MHz.

110/220 V
50-60 Hz


[Signature]

PULSE GENERATOR
PERFORMANCE CHECK

Model: **AVM-1-21 (POSITIVE CHANNEL)**

S.N.: **624**

Date: **SEPT 22 81**



- a) Output signal amplitude:
0 TO +5V
- b) Pulse width:
0 TO 6 μ SEC
- c) Rise time:
 \leq 100 μ SEC
- d) Fall time:
 \leq 135 μ SEC
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

5 V / DIV
1.0 μ SEC / DIV
4.0 MHz

110/220 V
50 - 60 Hz