


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: *AUH-5-1-05-P*

S.N.: *638*

Date: *NOV 13 81*



*10 VOLTS/DIV
1.0 NSEC/DIV
400 KHZ*

a) Output signal amplitude:

≥ 10 V

b) Pulse width:

≤ 135 PSECS


c) Jitter, stability:

OK

d) Prime power:

+15 ± 1 VOLT

190 mA



*10 VOLTS/DIV
100 PSEC/DIV
400 KHZ*

A handwritten signature or set of initials, possibly 'LJ', written in dark ink.