


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: *AVH-MPT*

S.N.: *644*

Date: *OCT 14 81*



1) Output signal amplitude:

*> 30 VOLTS*

2) Pulse width:

*< 1.0 NSEC*

3) Jitter, stability:

*OK*

4) Prime power:

*+15 ± 1.0 VOLT,*

*130 MA (MAX)*

*2.0 NSEC/DIV*

*32 VOLTS/DIV*

*400 KHZ*