


PULSE GENERATOR
PERFORMANCE CHECK

Model: **AVK-A-N**
S.N.: **658**
Date: **NOV 20 81**



- a) Output signal amplitude:
0.6 - 15V
- b) Pulse width:
3 TO 100 nSEC
- c) Rise time:
≤ 100 pSEC
- d) Fall time:
≤ 2 nSEC
- e) PRF:
SEE FILE 2
- f) Jitter, stability:
OK
- g) Prime power:

1.0 nSEC/DIV.
10 VOLTS/DIV
40 KHZ.

+15 ± 1.0 V.
225 mA