


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVM-1-D-PS-1P*

S.N.: *660*

Date: *NOV 30 81*



a) Output signal amplitude:

0 TO +5V

b) Pulse width:

0 TO 6 NSEC

c) Rise time:

≤ 100 PSEC

d) Fall time:

≤ 135 PSEC

e) PRF:

0 TO 25 MHz

f) Jitter, stability:

OK

g) Prime power:

120/240 V,

50 - 60 Hz

SWITCHABLE.

*5 VOLTS/DIV
1.0 NSEC/DIV
4.0 MHz.*