


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: **AVH-P**

S.N.: **675**

Date: **NOV 30 81**



a) Output signal amplitude:

$\geq 30 \text{ V}$

b) Pulse width:

$\leq 1.0 \text{ nsec}$

c) Jitter, stability:

OK

d) Prime power:

$\pm 15 \pm 1.0 \text{ VOLT.}$

130 mA.

16 VOLTS/DIV

1.0 nsec/DIV

400 KHz.