


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVP-AV-1-N-05-TTL-HV*
S.N.: *679*
Date: *NOV 6 81*



10 VOLTS/DIV
1.0 NS/CM/DIV
400 KHZ.

- a) Output signal amplitude: *0 TO -15 V*
- b) Pulse width* *0 TO 2 NS/CM*
- c) Rise time: *≤ 100 PS/CM*
- d) Fall time: *≤ 135 PS/CM*
- e) PRF: *0 TO 1.0 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *+15 ± 1.0 VOLT*

300 mA (MAX)

** NON-STANDARD
VALUES (-HV).*