


PULSE GENERATOR
PERFORMANCE CHECK

Model: **AUP-AU-1-N-05-TL**

S.N.: **692**

Date: **NOV 6 81**



- a) Output signal amplitude:
0 TO -10V
- b) Pulse width:
0 TO 4 NS
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:

10 VOLTS/DIV
1.0 NS/DIV
400 KHZ

+15 ± 1.0V
310 mA (MAX)

[Handwritten signature]