


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVM-1-P-EW-1*

S.N.: *696*

Date: *DEC 10 81*



a) Output signal amplitude:

*0 TO +5V*

b) Pulse width:

*0 TO 6 NSEC*

c) Rise time:

*≤ 100 PSEC*

d) Fall time:

*≤ 135 PSEC*

e) PRF:

*0 TO 25 MHz*

f) Jitter, stability:

*OK*

g) Prime power:

*5 VOLTS/DIV*

*1.0 NSEC/DIV*

*1.0 MHz*

*+24 ± 1.0 VOLTS*

*370 mA MAX.*

