


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVP-MV-1-N-14U-03-TL*

S.N.: *707*

Date: *DEC 23 81*



- a) Output signal amplitude:  
*0 TO 15 VOLTS*
- b) Pulse width:  
*0 TO 2 NS*
- c) Rise time:  
*≤ 100 PSEC*
- d) Fall time:  
*≤ 135 PSEC*
- e) PRF:  
*0 TO 1.0 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*10 VOLTS/DIV*

*1.0 NS/DIV*

*400 KHz.*

*+15 ± 1.0 VOLTS*

*265 mA*