


PULSE GENERATOR  
PERFORMANCE CHECK


Model: *AVN-S-2-P*

S.N.: *720*

Date: *JAN 14 82*



*5 V / DIV*  
*1.0 nSEC / DIV*  
*200 MHz*



*5 V / DIV*  
*1.0 nSEC / DIV*  
*250 MHz*

- a) Output signal amplitude:  
*0 TO +2 V*
- b) Pulse width:  
*300 TO 1500 pSEC*
- c) Rise time:  
*≤ 100 pSEC*
- d) Fall time:  
*≤ 100 pSEC*
- e) PRF:  
*50 TO 260 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*+24 ± 1.0 V*  
*325 mA (MAX)*