


PULSE GENERATOR

PERFORMANCE CHECK

Model: **AVL-2A-N-W-PS**

S.N.: **724**

Date: **JAN 29 82**



- a) Output signal amplitude:
-50 TO -180V
- b) Pulse width:
0 TO 400 NSEC
- c) Rise time:
≤ 1 NSEC
- d) Fall time:
≤ 2 NSEC.
- e) PRF:
0 TO 5 KHZ.
- f) Jitter, stability:
OK
- g) Prime power: .

100 V/DIV
10 NSEC/DIV
5 KHZ
R_i = 50 Ω

110/220 V
50-60 HZ.