


PULSE GENERATOR
PERFORMANCE CHECK

Model: **AVI-V-10**
S.N.: **736**
Date: **JAN 29 82**



- a) Output signal amplitude:
7, + 50 V.
- b) Pulse width:
0 TO 100 nsec.
- c) Rise time:
≤ 0.5 nsec
- d) Fall time:
≤ 2 nsec
- e) PRF:
0 TO 20 kHz
- f) Jitter, stability:
OK.
- g) Prime power:

32 V/DIV
10 nsec/DIV
20 kHz.
R_L = 50Ω.

+ 15 ± 1.0 V.
200 mA (max)
