


PULSE GENERATOR
PERFORMANCE CHECK

Model: **AVL-2-05-17**

S.N.: **739**

Date: **FEB 9 82**



- a) Output signal amplitude:
0 TO +350 V
- b) Pulse width:
0 TO 100 nSEC
- c) Rise time:
< 1 nSEC
- d) Fall time:
< 2 nSEC
- e) PRF:
0 TO 5 kHz
- f) Jitter, stability:
- g) Prime power: **OK.**

3.20 VOLTS/DIV.
10 nSEC/DIV.
3 kHz.

+15 ± 1.0 VOLTS

475 mA

WMA