


PULSE GENERATOR

PERFORMANCE CHECK

Model: **A20-2-P**

S.N.: **745**

Date: **FEB 9 82**



- a) Output signal amplitude:
0 TO +10 AMPS
- b) Pulse width:
0 TO 15 NSEC
- c) Rise time:
≤ 1 NSEC
- d) Fall time:
≤ 1 NSEC
- e) PRF:
0 TO 20KHZ
- f) Jitter, stability:
OK.
- g) Prime power:

20 AMP/DIV
10 NSEC/DIV
20 KHZ

+15 ± 1.0 VOLTS
200 mA MAX