


PULSE GENERATOR  
PERFORMANCE CHECK

Model: **AVR-1-PW-Q-17**

S.N.: **752**

Date: **APRIL 12 82**



a) Output signal amplitude:

**0 TO +300V**

b) Pulse width:

**0.1 TO 5.0  $\mu$ sec.**

c) Rise time:

**$\approx 15$  nsec**

d) Fall time:

**$\approx 15$  nsec**

e) PRF:

**0 TO 1 KHz**

f) Jitter, stability:

**OK**

g) Prime power:

**110/220V**

**50-60 Hz**

**50 V/DIV**

**0.1  $\mu$ sec/DIV**

**$R_L = 50 \Omega$**

**LS.**