


PULSE GENERATOR

PERFORMANCE CHECK


Model: **AVP-AV-2-N-TTL.**

S.N.: **764**

Date: **MAR 31 82**



**2 nSEC/DIV
10 VOLTS/DIV
50 KHZ**



**10 nSEC/DIV
10 VOLTS/DIV
50 KHZ.**

- a) Output signal amplitude:
0 TO -10 V
- b) Pulse width:
2 TO 50 nSEC
- c) Rise time:
≤ 100 pSEC
- d) Fall time:
≤ 135 pSEC
- e) PRF:
0 TO 50 KHZ
- f) Jitter, stability:
OK
- g) Prime power:

**+15 ± 1.0 V
225 mA.**