

IMPULSE GENERATOR

PERFORMANCE CHECK

Model:


AVH-5-N

S.N.:

784

Date:

FEB 24 82



a) Output signal amplitude:

$\approx 15 \text{ V}$

b) Pulse width:

$\leq 500 \text{ PSEC}$

c) Jitter, stability:

OK

d) Prime power:

$+15 \pm 1.0 \text{ VOLT}$

$180 \text{ mA}$

10 VOLTS/DIV

1.0 NSEC/DIV

400 KHZ