


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVH-M2-N*
S.N.: *451*
Date: *MAY 24 82*



- a) Output signal amplitude:
≥ -60 V
- b) Pulse width:
≤ 2 NS
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:

32 VOLTS/DIV
2 NS/DIV
400 KHz

+135 ± 5 VOLTS
5 mA (MAX)