


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO*
S.N.: *857*
Date: *APRIL 30 82*



- a) Output signal amplitude:
+10 TO +50 AMPS
- b) Pulse width:
20 ± 5 NSEC
- c) Rise time: *(HALF POWER)*
N/A
- d) Fall time:
N/A
- e) PRF:
0 TO 1 KHZ
- f) Jitter, stability:
OK
- g) Prime power:

25 AMP / DIV.
10 NSEC / DIV.
1 KHZ.

+15 ± 1.0 VOLT.
150 mA (MAX)
[Signature]