

PULSE GENERATOR

PERFORMANCE CHECK

Model:


AUL-2-N-PS

S.N.:

863

Date:

JUNE 23 82



a) Output signal amplitude:

0 TO 350V

b) Pulse width:

0 TO 100 NSEC

c) Rise time:

≈ 1.5 NSEC

d) Fall time:

≈ 2 NSEC

e) PRF:

0 TO 5 KHZ

f) Jitter, stability:

OK

g) Prime power:

110/220V,  
50 - 60 HZ

320V/CM  
10 NSEC/DIV  
2 KHZ.

44