

PULSE GENERATOR

PERFORMANCE CHECK

Model:


*AUL-2-N-PS*

S.N.:

*865*

Date:

*JUNE 23 82*



a) Output signal amplitude:

*0 TO -350V.*

b) Pulse width:

*0 TO 100 NSOR.*

c) Rise time:

*≈ 1.5 NSOR*

d) Fall time:

*≈ 2 NSOR*

e) PRF:

*0 TO 5 KHZ*

f) Jitter, stability:

*OK*

g) Prime power:

*320 V/DIV*

*10 NSOR/DIV*

*2 KHZ.*

*110/220V*

*50-60 Hz*