

PULSE GENERATOR

PERFORMANCE CHECK

Model:


17VI-U-N-A-PS

S.N.:

866

Date:

JUNE 23 82



32 V/DIV
10 nSEC/DIV
10 KHz

a) Output signal amplitude:

0 TO -50V

b) Pulse width:

0 TO 100 nSEC

c) Rise time:

≤ 0.5 nSEC

d) Fall time:

≤ 2 nSEC

e) PRF:

0 TO 20 KHz


f) Jitter, stability:

OK

g) Prime power:

110/220V

50-60 Hz



RISE TIME
32 V/DIV
1.0 nSEC/DIV
10 KHz.

[Handwritten signature]