


PULSE GENERATOR

PERFORMANCE CHECK

Model: **AVO-3B-P**

S.N.: **880**

Date: **JUNE 28 82**



- a) Output signal amplitude: **10 TO 30 AMPS**
- b) Pulse width: **1.5 NSEC**
- c) Rise time: **NA**
- d) Fall time: **NA**
- e) PRF: **0 TO 20 KHz**
- f) Jitter, stability: **OK**
- g) Prime power:

≈ 25 AMPS/DIV

2.0 NSEC/DIV

20 KHz

+15 ± 1.0 V

150 mA

(MAX)

[Handwritten Signature]