


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *ANM-1-P-D*

S.N.: *888*

Date: *MAY 31 82*



- a) Output signal amplitude:  
*0 TO +5V*
- b) Pulse width:  
*0 TO 6 NSEC*
- c) Rise time:  
*≤ 100 PSEC*
- d) Fall time:  
*≤ 150 PSEC*
- e) PRF:  
*0 TO 25 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*5 VOLTS/DIV*  
*1.0 NSEC/DIV*  
*4.0 MHz*

*+24 ± 1.0 VOLTS*  
*325 mA MAX*

*h) DELAY*  
*VARIABLE*  
*0 TO 5 NSEC.*

