


PULSE GENERATOR

PERFORMANCE CHECK


Model: *AVP-AU-HV2-05-N-M*

S.N.: *890*

Date: *MAY 24 82*



② *10 VOLTS/DIV*  
*1.0 NSEC/DIV*  
*400 KHZ*



⑥ *MONITOR OUTPUT*  
*PER ② ABOVE.*  
*SAME SCALES.*

- a) Output signal amplitude:  
*0 TO -20 V*
- b) Pulse width:  
*0 TO 2 NSEC*
- c) Rise time:  
*< 100 PSEC*
- d) Fall time:  
*< 200 PSEC*
- e) PRF:  
*0 TO 1.0 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*+15 ± 1.0 VOLT*

*275 mW*

- h) *MONITOR OUTPUT:*  
*20 db DOWN WRT*  
*MAIN OUTPUT (LESS*  
*DC OFFSET)*