


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVP-AU-1-N*

S.N.: *895*

Date: *MAY 24 82*



*10 VOLTS/DIV*  
*1.0 NSEC/DIV.*  
*400 KHz.*

a) Output signal amplitude:

*0 TO -10V*

b) Pulse width:

*0 TO 4 NSEC*

c) Rise time:

*≤ 100 PSEC*

d) Fall time:

*≤ 135 PSEC*

e) PRF:

f) Jitter, stability:

g) Prime power:

*+15 ± 1.0 VOLT.*

*300 mA*