


PULSE GENERATOR

PERFORMANCE CHECK

Model: **AVR-1-PW-05-P-250V**

S.N.: **898**

Date: **JUN 16 82**



- a) Output signal amplitude:
0 TO +250V
- b) Pulse width:
280 NS
- c) Rise time:
5 NS
≤ 10 NS
- d) Fall time:
≤ 20 NS
- e) PRF:
0 TO 1 KHz
- f) Jitter, stability:
OK
- g) Prime power:

110/220V
50 - 60 Hz.

50 V/DIV
100 NS/DIV
1 KHz
R_L = 50 Ω