


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-2-P*

S.N.: *901*

Date: *July 28 82*



*10 VOLTS/DIV
1.0 NS/DIV
4.0 MHz.*

- a) Output signal amplitude:
0 TO +15V
- b) Pulse width:
0 TO 2 NS
- c) Rise time:
≤ 100 PS
- d) Fall time:
≤ 135 PS
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

*+24 ± 1.0 V
450 mA (MAX)*

[Signature]