


PULSE GENERATOR

PERFORMANCE CHECK

Model: **AVO-2-P**

S.N.: **921**

Date: **JUNE 28 82**



- a) Output signal amplitude:
0 TO ± 10 AMP
- b) Pulse width:
0 TO 15 NSEC
- c) Rise time:
≤ 1 NSEC
- d) Fall time:
≤ 1 NSEC
- e) PRF:
0 TO 20 KHZ
- f) Jitter, stability:
OK
- g) Prime power:
+ 15 ± 1.0 V
250 mA.

6.7 AMP/DIV
10 NSEC/DIV
20 KHZ.