

IMPULSE GENERATOR

PERFORMANCE CHECK

Model:


AV11-MPT-N

S.N.:

932

Date:

JUN 31 82



a) Output signal amplitude:

≤ 30 VOLTS

b) Pulse width:

≤ 1 NSEC

c) Jitter, stability:

OK

d) Prime power:

$+15 \pm 1.0$ V

150 mA.

16 VOLTS/DIV

1.0 NSEC/DIV

400 KHz

A handwritten signature in black ink, appearing to be 'J. J. J.'