

IMPULSE GENERATOR

PERFORMANCE CHECK

Model:


1924 MP - N

S.N.:

933

Date:

Jul 31 82



a) Output signal amplitude:

≤ -30 VOLTS

b) Pulse width:

≤ 1 NSEC

c) Jitter, stability:

OK

d) Prime power:

$+15 \pm 1.0$ VOLT.

150 mA.

16 V VOLTS/DIV

1.0 NSEC/DIV

400 KHZ.