


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: AVH IMPT-2-N

S.N.: 934

Date: JULY 31 82



a) Output signal amplitude:

$\leq -60V$

b) Pulse width:

$\leq 2 NS$

c) Jitter, stability:

OK

d) Prime power:

$+15 \pm 1.0V$

160 mA.

32 VOLTS/DIV

2 NS/DIV

400 KHZ.