


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: *ACT-MPT-2-N*

S.N.: *936*

Date: *Aug 31 82*



- a) Output signal amplitude:
6-60V
- b) Pulse width:
2 nsec
- c) Jitter, stability:
OK
- d) Prime power:
+15 ± 1.0 vdc
160 mA

[Signature]

32 VOLTS/DIV
2 nsec/DIV
400 kHz