


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: *AV4*

S.N.: *955*

Date: *APR 30 82*



a) Output signal amplitude:

$\geq +30V$

b) Pulse width:

$\leq 1 \mu s$

c) Jitter, stability:

OK

d) Prime power:

$+15 \pm 1.0 \text{ VOLT}$

150 mA

*16 VOLT/DIV.
1.0 NS/DIV
400 KHZ*