


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVL-TN-1*

S.N.: *963*

Date: *Aug 30 82*



*20 nsec/div.
100 V/div.
2 kHz..*

- a) Output signal amplitude:
-230 V
- b) Pulse width:
5 to 100 nsec
- c) Rise time:
≤ 2 nsec
- d) Fall time:
≤ 3 nsec
- e) PRF:
0 to 5 kHz
- f) Jitter, stability:
OK
- g) Prime power:
*120 / 240 V,
50 - 60 Hz*

*h) DC OFFSET:
+30 VOLT, 6
mA MAX.*

i) PROP DELAY: ≤ 100 nsec