


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: AVH-5-P

S.N.: 971

Date: AUG 22 82



a) Output signal amplitude:

$\geq +20V$

b) Pulse width:

$\leq 0.5 \text{ nsec}$

c) Jitter, stability:

OK

d) Prime power:

$+15 \pm 1 \text{ VOLT}$

175 mA (MAX)

16 VOLTS/DIV
1.0 nSEC/DIV
400 kHz