


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: AVH-M1-N-A

S.N.: 979

Date: AUG 22 82



a) Output signal amplitude:

$\approx +30 \text{ V (0 to } +30)$

b) Pulse width:

$\leq 1 \text{ nsec}$

c) Jitter, stability:

OK

d) Prime power:

$+135 \text{ V} \pm 5 \text{ V}$

$5 \text{ mA (MAX)}$

32 VOLTS/DIV

1.0 nsec/DIV

400 kHz