


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *MP-AV-1-P-05-PS*

S.N.: *993*

Date: *NOV 30 82*



- a) Output signal amplitude:  
*0 TO 110V*
- b) Pulse width:  
*0 TO 4 NS*
- c) Rise time:  
*≤ 100 PSEC*
- d) Fall time:  
*≤ 135 PSEC*
- e) PRF:  
*0 TO 1 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*10 V / DIV.  
100 NS / DIV  
400 KHz.*

*110 / 220 V  
50 - 60 Hz.*

*AA*