


PULSE GENERATOR

PERFORMANCE CHECK


Model: *ANL-AV-1-C-N*

S.N.: *6017*

Date: *SEPT 4 1991*



Ⓐ 60db ATTN
100 VOLT/DIV



Ⓑ AS Ⓐ BUT
1 NS/DIV
(RISE TIME)

- a) Output signal amplitude:
0 TO -100 VOLTS
- b) Pulse width:
3 TO 100 NS
- c) Rise time:
≤ 1 NS
- d) Fall time:
≤ 2
- e) PRF:
0 TO 5 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120 / 240 V
50-60 Hz

[Signature]