


PULSE GENERATOR

PERFORMANCE CHECK

Model: ANPP-1-C-N

S.N.: 6018

Date: SEPT 17 1991



a) Output signal amplitude:
0 TO -10 VOLTS

b) Pulse width:
0.2 TO 100 NS

c) Rise time:
≤ 100 PS


d) Fall time:
≤ 130 PS

e) PRF:
0 TO 1.0 MHz

f) Jitter, stability:
OK

g) Prime power:
120 / 240 V
50 - 60 Hz

PG A 40 db ATTEN
10 VOLTS/DIV



PG B 40 db ATTEN
10 VOLTS/DIV
PRF = 1.0 MHz